

Quality Analysis ...



Innovative Technologies

Satmagan

Satmagan 135 is used to measure the magnetite content of iron ore samples. Accuracy is better than 0.4%. Samples are placed in special vials supplied by the manufacturer. The vial is placed in the measuring hole of the Satmagan 135 that has been calibrated with standards of varying magnetite content. The % magnetite is then determined.